

Federal Communications Commission

Before the
Federal Communications Commission
Washington, D.C. 20554

In the Matter of)
)
Amendment of the Commission’s Rules to Provide)
Spectrum for the Operation of Medical Body Area) ET Docket No. 08-59
Networks)
)

ERRATUM

Released: July 3, 2012

By the Chief, Office of Engineering and Technology:

On May 24, 2012, the Commission released a First Report and Order and Further Notice of Proposed Rulemaking, FCC 12-54, in the above captioned proceeding. This Erratum amends Appendix B as indicated below:

- 1. In SUBPART E heading, “TECHICAL” is corrected to read as “TECHNICAL”.
- 2. Section 95.628, in paragraph (f)(1), replace “(h)(2) and (h)(3)” with “(f)(2) and (f)(3)” and in paragraph (f)(2) correct “set forth in (f)” to read as “set forth in (d)”.
- 3. Section 95.639, in paragraph (5), correct “95.628(h)” to read as “95.628(f)”.

This Erratum also amends Appendix C as indicated below:

- 4. In paragraph 15, the third sentence is corrected to read as follows:

“An MBAN may be operated anywhere that CB station operation is authorized under § 95.405, except in the 2360-2390 MHz band MBAN use is restricted to indoor operation within a health care facility registered with the MBAN coordinator, and an MBAN is not required to transmit a station identification announcement.”

FEDERAL COMMUNICATIONS COMMISSION

Julius P. Knapp
Chief, Office of Engineering and Technology